

E30

E30 water purification system produce ultrapure and pure water for laboratory needs. It is designed for maximum convenience of use and to have maximum features. Is it a system with high price/ performance ratio.

ORDERING INFORMATION

Model	Part number
E30 Trace	CE30-1001
E30 HPLC	CE30-1101
E30 Bio	CE30-1201



DESCRIPTION

	E30 Trace	E30 HPLC	E30 Bio
Water type	• ultrapure water (Grade 1) • pure water (Grade 2)	• ultrapure water (Grade 1) • pure water (Grade 2)	• ultrapure water (Grade 1) • pure water (Grade 2)
Application	• atomic absorption spectrometry • plasma optical emission spectrometry • other inorganic trace analysis	• chromatography • mass spectrometry • microbiology • molecular biology	highly sensitive biology applications
Display	colour LCD display		
Conductivity sensor	•	•	•
TOC Monitor	-	•	•
Volumetric dispensing	•	•	•
Connection to Flow point	•	•	•
Storage tank	required but not included		
Installation	installable on a laboratory bench		

SPECIFICATIONS

	E30 Trace	E30 HPLC	E30 Bio
Grade 1 water resistivity at 25 °C	18.2 MΩ x cm	18.2 MΩ x cm	18.2 MΩ x cm
Grade 1 water conductivity at 25 °C	0.055 µS/cm	0.055 µS/cm	0.055 µS/cm
Grade 2 water conductivity at 25 °C	<0.1 µS/cm	<0.1 µS/cm	<0.1 µS/cm
Total Organic Carbon (TOC) level	<10 ppb	<5 ppb*	<5 ppb*
RNase	-	-	<0.01 ng/mL
DNase	-	-	<4 pg/µL
Bacteria	<0.01 CFU/mL	<0.01 CFU/mL	<0.01 CFU/mL
Endotoxins	<0.15 EU/mL	<0.15 EU/mL	<0.001 EU/mL
Particles >0.22 µm	<1 per mL	<1 per mL	<0.05/mL
Nominal flow, pure water (to storage tank)	10 L/h	10 L/h	10 L/h
Dimensions (WxDxH), cm	40x35x55	40x35x55	40x35x55
System weight, kg	17	18	19
Operation weight, kg	24	25	26
Feed water pressure	0.4- 6 bar	0.4- 6 bar	0.4- 6 bar
Feed water conductivity	< 1500 µS/cm	< 1500 µS/cm	< 1500 µS/cm

* In appropriate operating conditions <2 ppb, otherwise normally <5 ppb.

CONSUMABLES

Part number	Description	Replacement criteria	Comments
10320	Replacement pre-filter	If the filters are clogged or every 6 months	
10310	Replacement deionization module	Grade 2 water conductivity is > 0.5 µm/cm constantly or every 12 months	
10030	Polishing module "Polishing+"	Grade 1 water conductivity is > 0.1 µm/cm constantly or every 12 months	
10017	Replacement sterilization UV bulb	When required (on average every 2 years)	„Bio“ systems only
10018	Replacement photooxidation UV bulb	2 years on average	„HPLC“ and „Bio“ systems only
10012	Replacement 0.22 µm dispense microfilter	Every 6–12 months	„Trace“ and „HPLC“ systems
10120	Replacement ultrafilter	Every 3–6 months	„Bio“ systems only

FLOW DIAGRAMS

