

 Page 1/1	CERTIFICATE OF ANALYSIS CHROMagar™ Acinetobacter Base ref: AC092(B)	FR-LAB-197
		Version 9
		21-Aug-2024

Date of control: 19 NOV 2024	Lot number: P003734 Expiry date: OCT 2027
--	--

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature	7.0 +/- 0.2
after pouring and cooling at room temperature	
medium colour	yellow to pinkish
medium aspect	homogeneous

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum	absence of colonies
-------------	---------------------

Incubation for 24 h at 37 °C:

Antibacterial effect, by isolation method :

E. faecalis ATCC® 29212	no growth
-------------------------	-----------

Diameter, colour and aspect of isolated colonies :

A. baumannii AR5620	3-4 mm & red
A. baumannii AR5610	1-2 mm & red

This lot number satisfies QC test indicated as above.
Our QC should not be substituted for the user's own quality control.
For the preparation of media, please refer to the IFU

Signature and stamp :

