

Certificate of Analysis

Product Name: SARS-CoV-2/Influenza A+B/RSV Antigen Combo

Rapid Test (Nasopharyngeal Swab)

Catalog No.: ISIR-ACO535

Batch No.: 0058APR25IACO

Quantity: 1100 PCS

Expiry Date: 03/2027

Date of Sampling: 2025-04-11

Date of Analysis: 2025-04-11

Other information:

Buffer Lot: 0028APR25BACO, EXP: 03/2027

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
FLU A	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
FLU B	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
RSV	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
COV	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/
---------------	---

Final QC Conclusion:	This batch of product met the QC Criteria.
-----------------------------	--

QC supervisor: *Freeman zheng*

Date: 2025.04.11

