

Certificate of Analysis

Product Name: SARS-CoV-2/Influenza A+B/RSV Antigen Combo

Rapid Test (Nasopharyngeal Swab)

Catalog No.: ISIR-535

Batch No.: 0253FEB25IACO

Quantity: 1000PCS

Expiry Date: 01/2027

Date of Sampling: 2025-02-26

Date of Analysis: 2025-02-26

Other information:

Buffer Lot: 0058FEB25BACO, EXP: 01/2027

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
FLU A	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
FLU B	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
RSV	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
COV	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/
---------------	---

Final QC Conclusion:	This batch of product met the QC Criteria.
-----------------------------	--

QC supervisor: *Freeman. zheng*

Date: 2025.02.26

