

Page 1/1

Date of control:

CERTIFICATE OF ANALYSIS

CHROMagar Campylobacter base ref: CP572(B) / CP573-25(B)

FR-LAB-314

Version 5

29-Jan-2024

Lot number:

P003492

Expiry date:

MAY 2027

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature 7.4 +/- 0.2

after pouring and cooling at room temperature

02 JUL 2024

medium colour yellow

medium aspect homogeneous, translucide

MICROBIOLOGICAL PROPERTIES:

Sterility after minimum 3 days at 30 °C (medium preparation following IFU):

No inoculum absence of colonies

Incubation for 36 h at 42 °C:

Antibacterial effect, by isolation method:

E. faecalis ATCC® 29212 no growth K. oxytoca AR5755 no growth

Diameter, colour and aspect of isolated colonies:

C. coli CIP 70.80T 1 - 2 mm, red C. jejuni ATCC® 29428 0.4 - 0.5 mm, red

Chardine

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp:

CHROMagar 29 Avenue George Sand 93210 LA PLAINE ST-DENIS