

## Page 1/1

Date of control:

#### **CERTIFICATE OF ANALYSIS**

# CHROMagar™ Streptococcus base ref: CQ392(B)

FR-LAB-267

Version 6

27-Aug-2024

Lot number:

P003686

Expiry date:

**AUG-2027** 

#### **PHYSICO-CHEMICAL PROPERTIES:**

pH before heating at room temperature 7.3 +/- 0.2

After pouring and cooling at room temperature

medium colour grey

23-SEP-2024

medium aspect Homogeneous with precipitates

#### **MICROBIOLOGICAL PROPERTIES:**

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU):

No inoculum absence of colonies

### Incubation for 24 h at 37 °C:

Antibacterial effect, by isolation method:

E. coli ATCC® 25922 no growth

Diameter, colour and aspect of isolated colonies:

 S. agalactiae AR4186
 1 - 2 mm & light blue

 E. faecalis ATCC® 29212
 1 - 2 mm & mauve

 S. uberis ATCC® 700407
 1 - 1.5 mm & steel blue

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp:

CHROMagar 29 Avenue George Sand 210 LA PLAINE ST-DENIS