

 Page 1/1	CERTIFICATE OF ANALYSIS CHROMagar™ MRSA powder Ref: MR502	FR-LAB-092
		Version 14
		26-Aug-2024

Date of control:

07 JAN 2025

Lot number:

P003877

Expiry date:

DEC 2027

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature	6.9 +/- 0.2
after pouring and cooling at room temperature	
medium colour	yellow
medium aspect	homogeneous

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum	absence of colonies
-------------	---------------------

Incubation for 24 h at 37 °C

Antibacterial effect, by isolation method:

<i>E. coli</i> ATCC® 25922	no growth
<i>E. faecalis</i> ATCC® 29212	no growth
MSSA ATCC® 25923	no growth

Diameter, colour and aspect of isolated colonies :

MRSA AR4702	1-1.5 mm & mauve
-------------	------------------

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp :

