

Wuxi NEST Biotechnology Co., Ltd

Certificate of Analysis

Product Name	15 mL Centrifuge Tube, Bulk	Product No.	601002	Lot No.	102123CG01
DOM	2023-10-21	Expiration Date	2026-09		
No.	Item	Inspection items/basis			Result
1	Appearance	No missing materials, cracks, bubbles, pulling marks, etc			Pass
2	Size	Comforming to the blueprint			Pass
3	Centrifugal force test	Centrifugal test products have no cracking, no damage, no leakage			Pass
4	Sealing test	After fitting the tube body and tube cover, negative pressure test, no water, no bubbling			Pass
5	Freezing test	Freezing test product no freezing crack, no leakage			Pass
6	Packaging	Correct packaging materials and quantity; intact packaging			Pass
7	Sterilization	Red colored irradiation tag with Certificate of Irradiation			Pass
8	*Endotoxin Detection	≤0.05EU/ml			Pass
9	*Rnase test	No Rnase detected			Pass
10	*Dnase test	No Dnase detected			Pass
11	*Sterility testing	No microorganisms detected			Pass
Conclusion		Pass			
Note: Only Pass or No pass is applicable in filling in 'Result'. Items marked with "*" are periodic testing items for representative products.					
Add: No.530, Xida Road, Meicun Industrial Park, Xinwu District, Wuxi		Tested by	Jiang Meiqi	Approved by	He Yun
Tel: (+86) 0510-68006788					
Fax: (+86) 0510-68006788		Date	2023-10-24	Date	2023-10-24
https://www.cell-nest.com					

NEST