

## Page 1/1

Date of control:

### **CERTIFICATE OF ANALYSIS**

# CHROMagar™ mSuperCARBA™ base ref: SC172(B) / SC173-25(B) / SC173-10kg(B)

FR-LAB-304

Version 6

26-Aug-2024

Lot number:

P004047

Expiry date:

FEB-2028

## **PHYSICO-CHEMICAL PROPERTIES:**

pH before heating at room temperature 7.2 +/- 0.2

After pouring and cooling at room temperature

01-APR-2025

medium colour yellow

medium aspect homogeneous

#### **MICROBIOLOGICAL PROPERTIES:**

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum absence of colonies

#### Incubation for 24 h at 37 °C:

Antibacterial effect, by isolation method:

E. faecalis ATCC® 29212 no growth K. pneumoniae AR5185 no growth

Diameter, colour and aspect of isolated colonies:

K. pneumoniae KPC AR5251E. coli OXA48 AR59943 mm & steel blue2 mm & pink

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp:

CHROMagar 29 Avenue George Sand 3210 LA PLAINE ST-DENIS