

Page 1/1

Date of control:

CERTIFICATE OF ANALYSIS CHROMagar™ Serratia base ref: SM302(B)

FR-LAB-374 Version 5

27-Aug-2024

Lot number:

P003448

Expiry date:

JAN-2028

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature 7.1 +/- 0.2

After pouring and cooling at room temperature

18-FEB-2025

medium colour yellow medium aspect homog

homogeneous

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum absence of colonies

Incubation for 18 - 24 h at 37 °C:

Antibacterial effect, by isolation method:

E. faecalis AR5199 no growth
E. coli AR6240 no growth

Diameter, colour and aspect of isolated colonies:

S.marcescens AR5569 2 - 3 mm & blue green to metallic blue

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp:

CHROMagar
29 Avenue George Sand
93210 LA PLAINE ST-DENIS
FRANCE