

	CERTIFICATE OF ANALYSIS CHROMagar™ MDR Selective suppl. ref: CR102/CR103-25	FR-LAB-199
		Version 7
		28-Feb-2024

Date of control: **18 FEB 2025**

Lot number: **169044**
Expiry date: **DEC 2026**

PHYSICO-CHEMICAL PROPERTIES:

powder colour	white
dissolution in sterile water	immediate
aspect after reconstitution	transparent

MICROBIOLOGICAL PROPERTIES:

Incubation for 24 h at 37 °C:

Antibacterial effect, by isolation method :

<i>A. baumannii</i> non MDR AR5610	no growth
------------------------------------	-----------

Diameter, colour and aspect of isolated colonies :

<i>A. baumannii</i> MDR AR5620	3-4 mm & red
--------------------------------	--------------

This lot number satisfies QC test indicated as above.
Our QC should not be substituted for the user's own quality control.
For the preparation of media, please refer to the IFU

Signature and stamp :

