

	CERTIFICATE OF ANALYSIS LIM RambaQUICK Strep B base ref: LB082(B)		FR-LAB-292
			Version 8
	Page 1/1		30-Aug-2024

Date of control:	03-DEC-2024	Lot number:	P003760
		Expiry date:	OCT-2027

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature	7.4 +/- 0.2
after boiling and cooling at room temperature	
broth colour	yellow
broth aspect	transparent

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU):

No inoculum	no growth
-------------	-----------

Incubation according to LIM RambaQUICK™ StrepB method:

<i>S. agalactiae</i> AR4186	good growth
<i>E. faecium</i> AR5700	partial inhibition
<i>E. coli</i> AR3740	no growth

This lot number satisfies QC test indicated as above.
Our QC should not be substituted for the user's own quality control.
For the preparation of media, please refer to the IFU

Signature and stamp:



CHROMagar
29 Avenue George Sand
93210 LA PLAINE ST-DENIS
FRANCE