CHROMagar™	CERTIFICATE OF ANALYSIS CHROMagar™ Orientation ref: RT412 / RT413-25 / RT413-10Kg	FR-LAB-029
		Version 11
Page 1/1		26-Aug-2024

Date of control: **02-APR-2025**

Lot number: **P003998**
Expiry date: **FEB-2028**

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature	7.0 +/- 0.2
after pouring and cooling at room temperature	
medium colour	yellow
medium aspect	homogeneous

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum	absence of colonies

Incubation for 24 h at 37 °C:
Diameter, colour and aspect of isolated colonies:

E. coli AR3741	2 - 3 mm, red
P. mirabilis AR3022	1.5 - 2 mm, brown halo & no swarming
E. faecalis AR3908	0.5 mm, turquoise blue
K. pneumoniae K3875	2-3 mm, steel blue
S. saprophyticus AR3882	0.5 mm, pink
S. aureus AR3916	1 mm, golden
S. agalactiae AR3938	< 0.5 mm, light blue

This lot number satisfies QC test indicated as above.
Our QC should not be substituted for the user's own quality control.
For the preparation of media, please refer to the IFU

Signature and stamp:

Clotilde Mersch
Quality Manager

CHROMagar
29 Avenue George Sand
93210 LA PLAINE ST-DENIS
FRANCE