

	CERTIFICATE OF ANALYSIS CHROMagar™ Salmonella Plus Base ref: SA162(B) / SA163-25(B)	FR-LAB-143
		Version 9
		27-Aug-2024

Date of control:

10 FEB 2025

Lot number:

P003834

Expiry date:

JAN 2028

PHYSICO-CHEMICAL PROPERTIES:

pH before heating at room temperature 7.5 +/- 0.2

After pouring and cooling at room temperature

medium colour

transparent

medium aspect

homogeneous

MICROBIOLOGICAL PROPERTIES:

Contamination check after minimum 3 days at 30 °C (medium preparation following IFU) :

No inoculum

absence of colonies

Incubation for 24 h at 37 °C:

Antibacterial effect:

S. saprophyticus ATCC® 15305

full inhibition

A. baumannii SN96-461

full inhibition

Colour of isolated colonies :

Salmonella spp AR4053

2 - 3 mm & mauve

Salmonella lactose plus AR3910

2 - 2.5 mm & mauve may be slight mauve on isolated colonies

Salmonella Typhi DP7

0.5 - 1 mm & mauve may be slight mauve on isolated colonies

E. coli AR3740

1 - 1.5 mm & colourless

C. freundii AR3134


1.5 - 2 mm turquoise blue

This lot number satisfies QC test indicated as above.

Our QC should not be substituted for the user's own quality control.

For the preparation of media, please refer to the IFU.

Signature and stamp :


CHROMagar
 29 Avenue George Sand
 93210 LA PLAINE ST-DENIS
 FRANCE

Manufacturer : CHROMagar 29 Avenue George Sand 93210 La Plaine Saint-Denis France

Email: CHROMagar@CHROMagar.com Tel +33 (0)1.45.48.05.05. Website: www.CHROMagar.com

CHROMagar is a trademark created by Dr A. Rambach

ATCC® is a registered trademark of the American Type Culture Collection