



Certificate of Sterility

Date: 28 April 2025
Product/Material Number: 7-C1VS-99D-00 / SERIE 300 23cm x 23cm / 9 x 9
Lot Number: 199859
Irradiation Run Number: 1013-63810A
Irradiation Dose Specified: 25 kGy to 40 kGy
Irradiation Dose Delivered: 27.4 kGy to 37.7 kGy
Sterility Assurance Level (SAL): 10^{-6}
Manufacturing Date: 04/09/2025
Expiration Date: 04/2027

This certification is provided as assurance that the above product has been dosimetrically released as sterile. Sterility Assurance Level (SAL) determined by test method ANSI/AAMI/ISO 11137. All aspects of the manufacturing process are within specifications and the specified minimum dose has been delivered to the product.


Angela Swanson
Quality Assurance Manager